



The Model T100 UV Fluorescence SO₂ Analyzer



The Model T100 SO₂ analyzer uses the proven UV fluorescence principle and advanced electronics to allow accurate, dependable, continuous measurements for ambient air quality, stack gas monitoring and other applications.

— With NumaView™ premium T Series software —

- Large, vivid, and durable color touchscreen display
- All other T Series instrument platform features
- Lifetime technical support by phone and email
- Standard two-year warranty



T100 Specifications

■ Ranges	Min: 0 - 50 ppb full scale Max: 0 - 20,000 ppb full scale (selectable, dual-range supported)
■ Measurement Units	ppb, ppm, µg/m ³ , mg/m ³ (selectable)
■ Zero Noise	< 0.2 ppb (RMS)
■ Span Noise	< 0.5% of reading (RMS) above 50 ppb
■ Lower Detectable Limit	< 0.4 ppb
■ Zero Drift	< 0.5 ppb/24 hours
■ Span Drift	< 0.5% of full scale/24 hours
■ Lag Time	20 seconds
■ Rise/Fall Time	< 100 seconds to 95%
■ Linearity	1% of full scale
■ Precision	0.5% of reading above 50 ppb
■ Sample Flow Rate	650 cc/min ±10%
■ Power Requirements	100V-120V, 220V-240V, 50/60 Hz
■ Analog Output Ranges	10V, 5V, 1V, 0.1V (selectable)
■ Recorder Offset	±10%
■ Included I/O	1 x Ethernet: 10/100Base-T 2 x RS232 (300-115,200 baud) 2 x USB device ports 8 x opto-isolated digital outputs 6 x opto-isolated digital inputs 4 x analog outputs
■ Optional I/O	1 x USB com port 1 x RS485 8 x analog inputs (0-10V, 12-bit) 4 x digital alarm outputs Multidrop RS232 3 x 4 - 20mA current outputs
■ Operating Temperature Range	5 - 40°C (with US EPA Approval)
■ Dimensions (HxWxD)	7" x 17" x 23.5" (178 x 432 x 597 mm)
■ Weight	35.7 lbs (16.2 kg)
■ Certifications	US EPA: EQSA-0495-100 EU: EN14212 TÜV Rheinland QAL1 Certified: EN15267 MCerts: Sira MC0500067/07 CNEMC: 质(认)字 No. 2015-028 Report

Specifications subject to change without notice.
All specifications are based on constant conditions.



9970 Carroll Canyon Road ■ San Diego, CA 92131
Ph. 858-657-9800 Fax 858-657-9816
Email api-sales@teledyne.com

For more information about the Teledyne API family of monitoring instrumentation products, call us or visit our website at:

www.teledyne-api.com

© 2018 Teledyne API
Printed documents are uncontrolled. SAL000039G (DCN 7894) 05.01.18

